

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130290

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

S. Dunn #1361
ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

NOVEMBER 17, 2009

OPERATOR'S NAME:

DUNN/S

OPERATOR'S NUMBER: 4186

SUBJECT'S LAST NAME:

VERIFICATION OF CAL

SUBJECT'S FIRST NAME/MI:

V

O.L. #: V

DEPT/AGENCY: ANCI

CASE/REPORT: V

TEST TYPE: V

ALCO TARGET VALUE: .002

ALCO S/N: 77621

--- BREATH ANALYSIS ---

.002 ADJUSTED FOR 29.54 in
ALCO TARGET .000 01:33
BLANK TEST .000 01:33
INTERNAL STANDARD VERIFIED 01:34
ALCO TV 29.54 in .001 01:34
BLANK TEST .000 01:35
SUBJECT SAMPLE .000 01:35
BLANK TEST .000 01:36
ALCO TV 29.54 in .001 01:36
BLANK TEST .000 01:37

DIAGNOSTIC CHECK

S. Dunn #1361
ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130290

NOVEMBER 17, 2009

TIME 01:31

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 01/08/09
HEATERS
SAMPLE CHAMBER: 49c
BREATH TUBE: 41c
BAROMETER: 29.54 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./0123456789:;<=>?@ABCDEF
HIJKLMNOPQRSTUVWXYZ[]_`abdefghijklmnop
qrstuvwxyz{|}~

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument
State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

NOV 20 2009

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130290

Name STEVE DUNN

ID# 4106 Date 11-17-09

Agency APD

Phone # 764-1361

Instrument Location ANCHORAGE JAIL

B Alco S/N 77621 Target Value .082 High Pressure 150 PSI

Alco Test Value Average .081
1st Alco

.081
2nd Alco

Signature [Signature]

(OVER)

[Signature]
11/20/09

I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

[Signature]
Nita J. Bolz

Scientific Director

State Breath Alcohol Program

Subscribed and sworn before me this 18 day of Dec, 2009.

[Signature]
Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office

